

Technical Files Clinical Report (Saliva)	Part A/B: Part A
	File number:NRM-CE-142
	Revision:A/4

Nanjing Norman Biological Technology Co.,Ltd

Novel Coronavirus (2019-nCoV) Antigen

Testing Kit (Colloidal Gold)

(Saliva)

Clinical Report

Technical Files Clinical Report (Saliva)	Part A/B: Part A
	File number:NRM-CE-142
	Revision:A/4

Norman Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold)

(Saliva)

Clinical Study Protocol

Contents

- A. Trial Objectives
- B. Administrative Information
- C. Study Design
 - 1) Time Scale of Evaluation
 - 2) Materials used
 - 3) Specimens to be tested
 - 4) Data Storage and Reporting

Technical Files Clinical Report (Saliva)	Part A/B: Part A
	File number:NRM-CE-142
	Revision:A/4

1. Trial Objectives

The purpose of this evaluation is to establish the clinical specimens performance of the Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold), including performance on clinical saliva specimens. The data obtained will be used in the application for CE certification.

2. Administrative Information

- 1) The protocol should be read carefully, the protocol and 'Test Procedure' supplied with the reagents must be followed exactly unless explicitly stated.
- 2) The assays during the period of the evaluation should be double checked by another technician.
- 3) The results for both Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold) and any alternative assay methods must be properly identified. The data should be clearly legible and marked with the name of the laboratory and initialed by Coordinator.
- 4) All original raw data must be made available for further information.

3. Study Design

3.1 Random blinding requirements: According rules of Random coding, Set blinding, and Unblind.

3.2 Time Scale of Evaluation

The evaluation should be completed within 12 weeks.

3.3 Materials Used

- 1) Target reagent: A lot of qualified Norman Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold) produced based on technological process and the outputted Standard Operating Procedure. Test result with no analyzer or reader.
- 2) Reference reagent: PCR, the comparator assay: Real-Time Fluorescent RT-PCR Kit for Detecting SARS-CoV-2 from BGI Genomics.

3.4 Enrollment criteria: Cases with fever, fatigue, nasal congestion, runny nose, sore throat or dry cough; Exclusion criteria: Samples that cannot be processed in parallel determination test due to sampling or human factors (samples contaminated during operation).

3.5 VTM type, as applicable, for the antigen test and PCR: Virus extraction tube for antigen, virus Sample preservation tube for PCR.

3.6 Specimens collection and testing site: Infectious diseases hospital of Xinjiang Autonomous Region

3.7 User training: In case of the Antigen Test kit operator, training will be conducted on sample collection, transportation and treatment to ensure the correct use of the extract R1, sample processing, sample loading and result reading. The relevant operation video will also provide reference.

The applicant should be trained in the use of test reagents, especially the standardized training of random blind method according to the relevant national requirements.

Technical Files Clinical Report (Saliva)	Part A/B: Part A
	File number:NRM-CE-142
	Revision:A/4

3.8 Specimens to be tested

All the specimens used in this evaluation were form Infectious diseases hospital of Xinjiang Autonomous Region, which is the Clinical Cooperation Unit of Norman.

1) Specimen types

Saliva should be used.

2) Specimen quantity

Specimens above should be collected with known PCR positive and known PCR negative patients' samples after.

3) Other kinds of samples (e.g., nasal, OP, NP, Saliva, etc.) from the same person plan to tested by Company's Kit, if cloud get enough samples. each of them, should not less than 30 for specimens consistency.

4) Specimen number:

All kinds of Positive samples should not less than 110.

All kinds of Negative samples should not less than 200.

Specimens consistency samples from same patient, should not less than 30, Saliva, Nasopharyngeal swab and Oropharyngeal swab.

3.9 Data analysis

The product of this research is a qualitative detection clinical trial. The detection results of the two methods will be summarized in the form of 2×2 spreadsheet, and the sensitivity (positive coincidence rate), specificity (negative coincidence rate), total coincidence rate and other indicators will be calculated based on this and its 95% confidence interval.

3.10 Storage and Reporting

All data will be filed both on hard copy and in electronically files. Data will be stored for at least 5 years. All laboratory results are strictly confidential. Copies of raw data were retained under the internal R&D department of Norman for future reference.

Technical Files Clinical Report (Saliva)	Part A/B: Part A
	File number: NRM-CE-142
	Revision: A/4

Norman Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold) (Saliva) Performance on clinical Study Report

1. Purpose

1.1 To study the performance of the Norman Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold) with saliva specimens.

1.2 To compare the consistency between Saliva, Nasopharyngeal swab and Oropharyngeal swab, which are tested by Company's Kit

2. Method and material.

Specimen collected in hospital by qualified medical personnel and coded with random number blindly. The Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold) Test produced by Norman is used to screen COVID-19 PCR positive patient Samples and COVID-19 PCR negative patient Samples. 343 saliva specimens collected according to proper procedure.

Specimens consistency samples from same patient, were 60 pairs, positive 30 pairs (CP1-30), negative 30 pairs (CN1-30), Saliva, Nasopharyngeal swab and Oropharyngeal swab.

VTM type: Virus extraction tube for antigen, virus Sample preservation tube for PCR.

The test kits (LOT NO. RD0516 was used for evaluation.

Random blinding requirements: According rules of Random coding, Set blinding, and Unblind.

3. The results table (random blind testing)

3.1 results table

Clinical of Saliva

A total of 343 saliva samples consisting of 124 positive, 219 negative saliva specimens, were considered evaluable in this study. The performance of the Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold) as compared to the RT-PCR comparator method are presented in the table below:

Performance against the Comparator Method (saliva)

Method		Results	PCR Test		Total Results
Norman	Novel		positive	Negative	

Technical Files Clinical Report (Saliva)	Part A/B: Part A
	File number:NRM-CE-142
	Revision:A/4

Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold)	positive	113	2	115
	Negative	11	217	228
Total Results		124	219	343

Relative Sensitivity:	113/124	91.13% (84.68%~95.49%)
Relative Specificity:	217/219	99.09% (96.74%~99.89%)
Accuracy:	330/343	96.21% (93.61%~97.97%)

* 95% Confidence Interval

3.2 The consistency of different samples

Sample NO.	Nasopharyngeal swab	Oropharyngeal swab	Saliva
CP1	+++	+++	+++
CP2	+	+	+
CP3	++	+	+
CP4	+	+	+
CP5	++	++	++
CP6	+	+	+
CP7	++	+	+
CP8	+	+	+
CP9	++	++	++
CP10	+	+	+
CP11	+++	++	++
CP12	+	+	+
CP13	++	++	++
CP14	++	+	+
CP15	+	+	+
CP16	+++	+++	+++
CP17	+	+	+
CP18	+++	+++	+++

Technical Files
Clinical Report (Saliva)

Part A/B: Part A

File number:NRM-CE-142

Revision:A/4

CP19	++	++	++
CP20	+++	++	++
CP21	+	+	+
CP 22	+	+	+
CP 23	+	+	+
CP24	+	+	+
CP25	+	+	+
CP26	+	+	+
CP27	+	+	+
CP28	+	+	+
CP29	+	+	+
CP30	+	+	+
CN1	-	-	-
CN2	-	-	-
CN3	-	-	-
CN4	-	-	-
CN5	-	-	-
CN6	-	-	-
CN7	-	-	-
CN8	-	-	-
CN9	-	-	-
CN10	-	-	-
CN11	-	-	-
CN12	-	-	-
CN13	-	-	-
CN14	-	-	-
CN15	-	-	-
CN16	-	-	-

Technical Files Clinical Report (Saliva)	Part A/B: Part A
	File number:NRM-CE-142
	Revision:A/4

CN17	-	-	-
CN18	-	-	-
CN19	-	-	-
CN20	-	-	-
CN21	-	-	-
CN22	-	-	-
CN23	-	-	-
CN24	-	-	-
CN25	-	-	-
CN26	-	-	-
CN27	-	-	-
CN28	-	-	-
CN29	-	-	-
CN30	-	-	-

Result specification: “+”=Positive “-”=Negative

The consistency of different samples is good, Norman Novel Coronavirus (2019-nCoV) Antigen(Colloidal Gold) could be used for Saliva, Nasopharyngeal swab and Oropharyngeal swab samples.

4. Discussion and Conclusion

The results above showed that there was a good consistency between the Norman Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold) and the PCR results. And the results of different types of samples are consistent. So the Norman Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold) has a good clinical performance.

Principal Investigator(Reviewer)

Yao Wenhai

R&D Approver

Zhang Naiyuan

Date:2020-10-13